



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120821\
 Data File : VX025611.D
 Acq On : 08 Dec 2021 09:13
 Operator : JC/MD
 Sample : BFB620
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB620

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM112221WMA.M
 Title : VOC Analysis
 Last Update : Sun Dec 12 06:39:29 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	4589	PASS
75	95	30	80	50.2	12262	PASS
95	95	100	100	100.0	24416	PASS
96	95	5	9	6.9	1681	PASS
173	174	0.00	2	0.9	177	PASS
174	95	50	120	79.5	19410	PASS
175	174	5	9	8.2	1591	PASS
176	174	95	101	98.7	19149	PASS
177	176	5	9	6.9	1318	PASS

m/z	Abundance
40.10	173.0
43.90	115.0

Instrument :
MSVOA_X
ClientSampleId :
BFB620

m/z	Abundance
36.00	136.0
37.10	916.0
38.10	933.0
39.10	341.0
39.90	155.0
44.10	302.0
44.90	192.0
46.90	168.0
48.00	114.0
49.10	846.0
50.00	3750.0
51.00	1094.0
55.00	63.0
56.00	356.0
57.00	600.0
58.00	50.0
60.10	131.0
61.00	952.0
62.00	883.0
63.00	782.0
67.00	53.0
68.00	1851.0
69.00	1845.0
70.00	194.0
72.00	135.0
73.00	768.0
74.00	3362.0
75.00	8956.0
76.00	820.0
77.00	80.0
78.00	54.0
78.90	714.0
79.90	185.0
80.90	756.0
81.80	187.0
87.00	746.0
88.00	753.0
90.80	62.0
92.00	487.0
93.00	688.0
94.00	2011.0
95.00	17360.0
96.00	1300.0
103.80	83.0
106.00	87.0
115.90	92.0
117.10	100.0
117.90	75.0
118.90	107.0
129.80	63.0
136.90	57.0
140.80	201.0
143.00	161.0
155.00	62.0
172.80	121.0
174.00	11397.0
175.00	1089.0
176.00	11334.0
177.00	746.0

Instrument :
MSVOA_X
ClientSampleId :
BFB620

m/z	Abundance
36.10	216.0
37.00	1420.0
38.10	1271.0
39.10	461.0
40.10	146.0
44.10	415.0
45.10	235.0
47.10	338.0
48.00	142.0
49.10	1275.0
50.00	5585.0
51.10	1823.0
55.10	85.0
56.00	528.0
57.00	935.0
60.00	276.0
61.10	1263.0
62.00	1171.0
63.00	1144.0
63.90	96.0
67.00	81.0
68.00	3266.0
69.00	2856.0
70.00	246.0
71.90	118.0
73.00	1215.0
74.00	4957.0
75.00	14904.0
76.00	1280.0
76.90	183.0
78.20	70.0
78.90	1142.0
79.90	318.0
80.90	1169.0
81.90	263.0
87.00	1193.0
88.00	1057.0
90.90	114.0
92.00	1023.0
93.00	1380.0
94.10	3600.0
95.00	28968.0
96.00	1980.0
103.90	174.0
105.80	151.0
115.80	93.0
116.90	215.0
117.90	163.0
118.90	144.0
127.90	111.0
130.00	114.0
130.90	65.0
136.80	65.0
140.80	330.0
143.00	323.0
148.00	56.0
172.90	218.0
174.00	23352.0
175.00	1864.0
176.00	22952.0
177.00	1642.0

Instrument :
MSVOA_X
ClientSampleId :
BFB620

m/z	Abundance
36.00	212.0
37.00	1109.0
38.10	1043.0
39.00	463.0
40.00	176.0
44.00	302.0
45.00	184.0
47.10	286.0
48.00	124.0
49.00	837.0
50.10	4433.0
51.10	1389.0
56.00	499.0
57.00	668.0
59.90	196.0
61.00	1198.0
62.00	1195.0
63.10	804.0
63.90	85.0
68.00	2521.0
69.00	2492.0
70.10	212.0
72.00	189.0
73.00	1141.0
74.00	4193.0
75.00	12927.0
76.00	1036.0
77.00	155.0
77.90	77.0
78.90	1082.0
79.90	344.0
80.90	1032.0
81.90	274.0
87.00	1059.0
88.00	1047.0
90.90	138.0
92.00	885.0
93.10	1078.0
94.00	3218.0
95.00	26920.0
96.00	1763.0
103.90	98.0
106.00	136.0
115.80	100.0
116.90	158.0
117.90	161.0
118.70	149.0
119.00	141.0
127.90	138.0
128.80	77.0
129.70	82.0
141.00	263.0
142.90	340.0
148.00	52.0
155.00	77.0
173.00	191.0
174.00	23480.0
175.00	1819.0
176.00	23160.0
177.00	1567.0

Instrument :
MSVOA_X
ClientSampleId :
BFB620