

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120822\
 Data File : VX033311.D
 Acq On : 08 Dec 2022 10:22
 Operator : JC/MD
 Sample : VX1208WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1208WBL01

Quant Time: Dec 09 00:42:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	245164	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95384	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48793	49.738	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.480%
35) Dibromofluoromethane	5.385	113	54158	49.114	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.220%
50) Toluene-d8	8.647	98	111440	48.310	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.620%
62) 4-Bromofluorobenzene	11.079	95	83860	47.545	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120822\
Data File : VX033311.D
Acq On : 08 Dec 2022 10:22
Operator : JC/MD
Sample : VX1208WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1208WBL01

Quant Time: Dec 09 00:42:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

