

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120822\
 Data File : VX033319.D
 Acq On : 08 Dec 2022 13:32
 Operator : JC/MD
 Sample : N5906-09DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-28DL

Quant Time: Dec 09 00:46:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	153826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	244174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213255	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47701	48.337	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.680%
35) Dibromofluoromethane	5.391	113	51198	46.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.240%
50) Toluene-d8	8.653	98	108391	47.178	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.360%
62) 4-Bromofluorobenzene	11.079	95	83606	47.593	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.180%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.489	96	18171	9.967	ug/l	90
44) Trichloroethene	7.129	130	11190	6.501	ug/l	98
64) Tetrachloroethene	9.275	164	57204	39.229	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120822\
Data File : VX033319.D
Acq On : 08 Dec 2022 13:32
Operator : JC/MD
Sample : N5906-09DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-28DL

Quant Time: Dec 09 00:46:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

