

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120822\
 Data File : VX033325.D
 Acq On : 08 Dec 2022 15:51
 Operator : JC/MD
 Sample : N5906-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB120622

Quant Time: Dec 09 00:49:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227762	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	44871	48.985	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.960%
35) Dibromofluoromethane	5.385	113	49052	47.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.760%
50) Toluene-d8	8.647	98	104214	48.629	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.260%
62) 4-Bromofluorobenzene	11.079	95	77733	47.438	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.880%
Target Compounds						
16) Acetone	2.398	43	1005	1.483	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120822\
Data File : VX033325.D
Acq On : 08 Dec 2022 15:51
Operator : JC/MD
Sample : N5906-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB120622

Quant Time: Dec 09 00:49:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

