

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120823\
 Data File : VX039100.D
 Acq On : 08 Dec 2023 12:51
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 08 23:47:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	9703	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	82130	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	99731	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	42370	31.919	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.400%
60) 4-Bromofluorobenzene	11.079	95	63638	29.214	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.367%
63) Toluene-d8	8.647	98	118095	29.950	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.833%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120823\
Data File : VX039100.D
Acq On : 08 Dec 2023 12:51
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 08 23:47:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 27 03:22:25 2023
Response via : Initial Calibration

