

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120823\
 Data File : VX039101.D
 Acq On : 08 Dec 2023 13:14
 Operator : JC/MD
 Sample : 05717-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-058-IDWGW-20231207

Quant Time: Dec 08 23:47:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2023
 Supervised By :Mahesh Dadoda 12/11/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	12268	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	100119	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	125454	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	52007	30.987	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 103.300%			
60) 4-Bromofluorobenzene	11.079	95	80709	29.454	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 98.167%			
63) Toluene-d8	8.647	98	146124	29.460	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 98.200%			
Target Compounds						Qvalue
15) Acetone	2.380	58	961	2.077	ug/l	# 57
25) Chloroform	5.087	83	2864m	0.814	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120823\
Data File : VX039101.D
Acq On : 08 Dec 2023 13:14
Operator : JC/MD
Sample : 05717-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-058-IDWGW-20231207

Quant Time: Dec 08 23:47:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 27 03:22:25 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2023
Supervised By :Mahesh Dadoda 12/11/2023

