

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120823\
 Data File : VX039102.D
 Acq On : 08 Dec 2023 13:37
 Operator : JC/MD
 Sample : 05717-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-059-IDWGW-20231207

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2023
 Supervised By :Mahesh Dadoda 12/11/2023

Quant Time: Dec 08 23:48:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	11077	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	93237	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	114178	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	48030	31.694	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	105.633%	
60) 4-Bromofluorobenzene	11.079	95	73562	29.497	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	98.333%	
63) Toluene-d8	8.646	98	135274	29.966	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	99.900%	
Target Compounds						Qvalue
15) Acetone	2.379	58	1073	2.568	ug/l	78
37) Trichloroethene	7.128	130	1505m	0.779	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120823\
Data File : VX039102.D
Acq On : 08 Dec 2023 13:37
Operator : JC/MD
Sample : 05717-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-059-IDWGW-20231207

Quant Time: Dec 08 23:48:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 27 03:22:25 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2023
Supervised By :Mahesh Dadoda 12/11/2023

