

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
 Data File : VX048750.D
 Acq On : 08 Dec 2025 13:35
 Operator : JC/MD
 Sample : Q3787-02MS
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15B-42.5-120425MS

Quant Time: Dec 09 04:09:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	137818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	254506	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	219870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	118869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	112523	60.831	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	121.660%#
35) Dibromofluoromethane	5.373	113	100776	57.508	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	115.020%
50) Toluene-d8	8.635	98	284538	46.322	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.640%
62) 4-Bromofluorobenzene	11.061	95	99977	47.200	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.400%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.386	62	2106	1.116	ug/l #	88
27) cis-1,2-Dichloroethene	4.477	96	5959	3.233	ug/l	95
44) Trichloroethene	7.117	130	12005	6.466	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
Data File : VX048750.D
Acq On : 08 Dec 2025 13:35
Operator : JC/MD
Sample : Q3787-02MS
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15B-42.5-120425MS

Quant Time: Dec 09 04:09:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

