

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
 Data File : VX048751.D
 Acq On : 08 Dec 2025 13:55
 Operator : JC/MD
 Sample : Q3787-03MSD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15B-42.5-120425MSD

Quant Time: Dec 09 04:10:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129588	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.744	114	275109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	225069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	122026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	108707	62.500	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	125.000%#
35) Dibromofluoromethane	5.373	113	81466	43.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.020%
50) Toluene-d8	8.634	98	232471	35.011	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	70.020%#
62) 4-Bromofluorobenzene	11.061	95	98901	43.195	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.400%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.380	62	1883	1.061	ug/l #	88
16) Acetone	2.385	43	2267	3.405	ug/l #	65
27) cis-1,2-Dichloroethene	4.489	96	6428	3.709	ug/l	97
44) Trichloroethene	7.116	130	9889	4.928	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
Data File : VX048751.D
Acq On : 08 Dec 2025 13:55
Operator : JC/MD
Sample : Q3787-03MSD
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15B-42.5-120425MSD

Quant Time: Dec 09 04:10:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

