

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
 Data File : VX048760.D
 Acq On : 08 Dec 2025 17:00
 Operator : JC/MD
 Sample : Q3788-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-01B-66.5-120525

Quant Time: Dec 09 04:13:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	95592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	193994	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	314319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	113748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	75815	59.091	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 118.180%#	
35) Dibromofluoromethane	5.379	113	65834	49.287	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.580%	
50) Toluene-d8	8.635	98	298903	63.839	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 127.680%#	
62) 4-Bromofluorobenzene	11.061	95	94284	58.397	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.800%	
Target Compounds						
						Qvalue
16) Acetone	2.374	43	7191	14.641	ug/l	97
44) Trichloroethene	7.104	130	3076	2.174	ug/l	82
64) Tetrachloroethene	9.263	164	1059	0.462	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
Data File : VX048760.D
Acq On : 08 Dec 2025 17:00
Operator : JC/MD
Sample : Q3788-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-01B-66.5-120525

Quant Time: Dec 09 04:13:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

