

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
 Data File : VX048761.D
 Acq On : 08 Dec 2025 17:21
 Operator : JC/MD
 Sample : Q3788-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-05-465-120525

Quant Time: Dec 09 04:13:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	158795	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	333603	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	259658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129696	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	136681	64.130	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 128.260%#	
35) Dibromofluoromethane	5.373	113	85386	37.173	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 74.340%#	
50) Toluene-d8	8.634	98	360181	44.734	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 89.460%#	
62) 4-Bromofluorobenzene	11.061	95	108254	38.990	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 77.980%#	
Target Compounds						Qvalue
16) Acetone	2.373	43	54679	67.016	ug/l	99
52) Toluene	8.702	92	18766	3.337	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
Data File : VX048761.D
Acq On : 08 Dec 2025 17:21
Operator : JC/MD
Sample : Q3788-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-05-465-120525

Quant Time: Dec 09 04:13:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

