

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX0137848.D
 Acq On : 09 Dec 2019 13:50
 Operator : JC/SP
 Sample : K6185-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 10 03:49:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	628096	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	958156	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	858026	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	372740	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	348605	47.81	ug/l	0.00
Spiked Amount			Recovery	=	95.62%	
35) Dibromofluoromethane	5.48	113	289055	48.80	ug/l	0.00
Spiked Amount			Recovery	=	97.60%	
50) Toluene-d8	8.71	98	1139799	49.39	ug/l	0.00
Spiked Amount			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	11.13	95	376803	45.28	ug/l	0.00
Spiked Amount			Recovery	=	90.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX0137848.D
Acq On : 09 Dec 2019 13:50
Operator : JC/SP
Sample : K6185-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 10 03:49:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

