

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX0137852.D
 Acq On : 09 Dec 2019 15:23
 Operator : JC/SP
 Sample : K6184-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 10 03:59:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	616214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	941664	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	831164	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	359705	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	344139	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
35) Dibromofluoromethane	5.49	113	278762	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
50) Toluene-d8	8.71	98	1110865	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	11.14	95	365939	44.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.48%	
Target Compounds						
24) 1,1-Dichloroethane	3.68	63	7368	0.658	ug/l #	92
27) cis-1,2-Dichloroethene	4.58	96	123946	16.894	ug/l	89
30) Chloroform	5.20	83	20571	1.811	ug/l	89
44) Trichloroethene	7.21	130	630308	88.217	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX0137852.D
Acq On : 09 Dec 2019 15:23
Operator : JC/SP
Sample : K6184-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 10 03:59:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

