

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX0137875.D
 Acq On : 10 Dec 2019 01:30
 Operator : JC/SP
 Sample : K6111-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 10 05:51:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	636623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	960783	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	872607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	386536	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	345207	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
35) Dibromofluoromethane	5.49	113	289423	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
50) Toluene-d8	8.71	98	1151019	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.14	95	385847	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
Target Compounds						
16) Acetone	2.44	43	61375	16.822	ug/l	99
20) Methylene Chloride	2.85	84	67457	9.624	ug/l	97
43) Isopropyl Acetate	6.43	43	132719	8.239	ug/l	99
95) Naphthalene	13.83	128	533433	20.373	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX0137875.D
Acq On : 10 Dec 2019 01:30
Operator : JC/SP
Sample : K6111-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 10 05:51:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

