

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX0137877.D
 Acq On : 10 Dec 2019 02:17
 Operator : JC/SP
 Sample : K6118-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 10 05:56:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	605138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	911224	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	831442	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	377792	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	333692	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
35) Dibromofluoromethane	5.49	113	272950	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	8.71	98	1091631	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.14	95	377595	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
Target Compounds						
16) Acetone	2.44	43	39029	11.254	ug/l	98
20) Methylene Chloride	2.85	84	1028898	154.429	ug/l	97
43) Isopropyl Acetate	6.43	43	133732	8.754	ug/l	99
95) Naphthalene	13.83	128	66530	2.600	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX0137877.D
Acq On : 10 Dec 2019 02:17
Operator : JC/SP
Sample : K6118-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 10 05:56:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

