

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX013847.D
 Acq On : 09 Dec 2019 13:21
 Operator : JC/SP
 Sample : VX1209WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1209WBL01

Quant Time: Dec 10 03:46:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	614301	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	941182	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	836596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	380177	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	341531	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	
35) Dibromofluoromethane	5.48	113	280625	48.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.46%	
50) Toluene-d8	8.71	98	1111585	49.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.06%	
62) 4-Bromofluorobenzene	11.13	95	382032	46.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.46%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX013847.D
Acq On : 09 Dec 2019 13:21
Operator : JC/SP
Sample : VX1209WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1209WBL01

Quant Time: Dec 10 03:46:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

