

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX013856.D
 Acq On : 09 Dec 2019 17:20
 Operator : JC/SP
 Sample : K6185-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20191204

Quant Time: Dec 10 04:20:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	619547	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	958952	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	850945	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	371475	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	349666	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
35) Dibromofluoromethane	5.48	113	285542	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	8.71	98	1124626	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
62) 4-Bromofluorobenzene	11.14	95	373709	44.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.74%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	2.37	101	31948	5.409	ug/l	98
12) 1,1-Dichloroethene	2.37	96	18910	3.207	ug/l	91
16) Acetone	2.43	43	7449	2.098	ug/l	94
27) cis-1,2-Dichloroethene	4.59	96	13860	1.879	ug/l	90
44) Trichloroethene	7.21	130	4699606	645.889	ug/l	97
49) 1,4-Dioxane	7.74	88	3436	19.651	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	3576	0.539	ug/l #	89
64) Tetrachloroethene	9.33	164	29323	4.621	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX013856.D
Acq On : 09 Dec 2019 17:20
Operator : JC/SP
Sample : K6185-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20191204

Quant Time: Dec 10 04:20:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

