

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX013860.D
 Acq On : 09 Dec 2019 18:54
 Operator : JC/SP
 Sample : K6185-06DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20191205DL

Quant Time: Dec 10 04:40:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	615446	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	946670	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	851679	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	366184	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	343596	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
35) Dibromofluoromethane	5.49	113	280885	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	8.71	98	1127474	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	11.14	95	371893	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	6320	1.077	ug/l	90
16) Acetone	2.44	43	10313	2.924	ug/l	92
27) cis-1,2-Dichloroethene	4.59	96	7484	1.021	ug/l	84
44) Trichloroethene	7.21	130	279728	38.943	ug/l	97
64) Tetrachloroethene	9.33	164	17511	2.757	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX013860.D
Acq On : 09 Dec 2019 18:54
Operator : JC/SP
Sample : K6185-06DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20191205DL

Quant Time: Dec 10 04:40:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

