

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120919\
 Data File : VX013862.D
 Acq On : 09 Dec 2019 19:41
 Operator : JC/SP
 Sample : K6185-10DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20191205DL

Quant Time: Dec 10 04:46:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	597113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	928661	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	814281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	354819	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	340490	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
35) Dibromofluoromethane	5.49	113	277114	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	8.71	98	1091760	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	11.14	95	355439	44.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.14%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	268701	47.199	ug/l	98
16) Acetone	2.43	43	11007	3.217	ug/l	90
44) Trichloroethene	7.21	130	33656	4.776	ug/l	99
64) Tetrachloroethene	9.34	164	8417	1.386	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120919\
Data File : VX013862.D
Acq On : 09 Dec 2019 19:41
Operator : JC/SP
Sample : K6185-10DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20191205DL

Quant Time: Dec 10 04:46:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

