


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX120919\
Data File : VX0137869.D
Acq On : 09 Dec 2019 22:47
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Title : SW846 8260
Last Update : Tue Nov 26 15:53:00 2019



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1674

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	23702	PASS
75	95	30	60	48.9	60587	PASS
95	95	100	100	100.0	123965	PASS
96	95	5	9	6.6	8222	PASS
173	174	0.00	2	1.5	1736	PASS
174	95	50	100	91.3	113141	PASS
175	174	5	9	7.5	8528	PASS
176	174	95	101	100.4	113593	PASS
177	176	5	9	6.6	7501	PASS