

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019773.D
 Acq On : 09 Dec 2020 13:04
 Operator : JC/MD
 Sample : L4965-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT012-1-3

Quant Time: Dec 10 05:56:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	347619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	567144	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	515778	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231935	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	210078	45.98	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.96%
35) Dibromofluoromethane	5.46	113	173226	47.26	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.52%
50) Toluene-d8	8.70	98	684642	48.93	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.86%
62) 4-Bromofluorobenzene	11.12	95	242722	45.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.14%

Target Compounds

					Qvalue
16) Acetone	2.42	43	17202	10.017 ug/l	99
20) Methylene Chloride	2.83	84	26754	5.560 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019773.D
Acq On : 09 Dec 2020 13:04
Operator : JC/MD
Sample : L4965-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
CRT012-1-3

Quant Time: Dec 10 05:56:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

