

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019779.D
 Acq On : 09 Dec 2020 15:24
 Operator : JC/MD
 Sample : L4971-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5S1-A-TOP

Quant Time: Dec 10 06:08:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	338559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	556167	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	518122	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	229844	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	209711	47.13	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.26%		
35) Dibromofluoromethane	5.46	113	169929	47.28	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.56%		
50) Toluene-d8	8.70	98	678087	49.41	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.82%		
62) 4-Bromofluorobenzene	11.12	95	239944	45.94	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.88%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	28597	17.098	ug/l 98
20) Methylene Chloride	2.84	84	22998	4.736	ug/l 99
43) Isopropyl Acetate	6.41	43	20343	2.466	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019779.D
Acq On : 09 Dec 2020 15:24
Operator : JC/MD
Sample : L4971-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5S1-A-TOP

Quant Time: Dec 10 06:08:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

