

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019781.D
 Acq On : 09 Dec 2020 16:11
 Operator : JC/MD
 Sample : L4971-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5S1-B-TOP

Quant Time: Dec 10 06:11:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	333478	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	536255	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	487534	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208331	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	203402	46.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.82%	
35) Dibromofluoromethane		5.46	113	164990	47.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.22%	
50) Toluene-d8		8.70	98	654157	49.44	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.88%	
62) 4-Bromofluorobenzene		11.12	95	225937	44.86	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	14463	8.779	ug/l	99
20) Methylene Chloride	2.83	84	11443	1.668	ug/l	98
43) Isopropyl Acetate	6.41	43	23728	2.983	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019781.D
Acq On : 09 Dec 2020 16:11
Operator : JC/MD
Sample : L4971-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5S1-B-TOP

Quant Time: Dec 10 06:11:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

