

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019782.D
 Acq On : 09 Dec 2020 16:34
 Operator : JC/MD
 Sample : L4971-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5S1-B-BOTTOM

Quant Time: Dec 10 06:12:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	334460	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	549316	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	507785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	225684	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	207487	47.20	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.40%	
35) Dibromofluoromethane		5.46	113	169888	47.85	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.70%	
50) Toluene-d8		8.70	98	670600	49.48	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.96%	
62) 4-Bromofluorobenzene		11.12	95	240164	46.56	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	13547	8.199	ug/l	100
20) Methylene Chloride	2.83	84	10944	1.522	ug/l	95
43) Isopropyl Acetate	6.41	43	27743	3.405	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019782.D
Acq On : 09 Dec 2020 16:34
Operator : JC/MD
Sample : L4971-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5S1-B-BOTTOM

Quant Time: Dec 10 06:12:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

