

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019791.D
 Acq On : 09 Dec 2020 20:03
 Operator : JC/MD
 Sample : L4971-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5S2-D-TOP

Quant Time: Dec 10 06:28:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	330621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	541035	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	489485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206302	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	204673	47.10	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.20%		
35) Dibromofluoromethane	5.46	113	166379	47.58	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.16%		
50) Toluene-d8	8.70	98	663905	49.73	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.46%		
62) 4-Bromofluorobenzene	11.12	95	224973	44.28	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.56%		

Target Compounds					Qvalue
16) Acetone	2.42	43	13617	8.337 ug/l	98
20) Methylene Chloride	2.84	84	10526	1.442 ug/l	96
43) Isopropyl Acetate	6.41	43	26925	3.355 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019791.D
Acq On : 09 Dec 2020 20:03
Operator : JC/MD
Sample : L4971-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5S2-D-TOP

Quant Time: Dec 10 06:28:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

