

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019796.D
 Acq On : 09 Dec 2020 23:32
 Operator : JC/MD
 Sample : VX1209WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1209WBL02

Quant Time: Dec 10 06:50:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	336721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	548498	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	496915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	220465	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	207514	46.89	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.78%
35) Dibromofluoromethane	5.46	113	167005	47.11	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.22%
50) Toluene-d8	8.70	98	666755	49.27	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.54%
62) 4-Bromofluorobenzene	11.12	95	232554	45.15	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.30%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019796.D
Acq On : 09 Dec 2020 23:32
Operator : JC/MD
Sample : VX1209WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX1209WBL02

Quant Time: Dec 10 06:50:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

