

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019799.D
 Acq On : 10 Dec 2020 00:42
 Operator : JC/MD
 Sample : L4993-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 10 06:55:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	335889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	545728	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	496264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	219624	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	204013	46.21	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.42%
35) Dibromofluoromethane	5.46	113	165468	46.92	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.84%
50) Toluene-d8	8.70	98	655736	48.70	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.40%
62) 4-Bromofluorobenzene	11.12	95	229176	44.72	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.44%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019799.D
Acq On : 10 Dec 2020 00:42
Operator : JC/MD
Sample : L4993-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 10 06:55:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

