

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019801.D
 Acq On : 10 Dec 2020 01:29
 Operator : JC/MD
 Sample : L4993-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TU032-32-6132-120720

Quant Time: Dec 10 07:03:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	333097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	543188	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	497330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	217227	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	205202	46.87	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.74%
35) Dibromofluoromethane	5.46	113	166087	47.31	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.62%
50) Toluene-d8	8.70	98	652870	48.71	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.42%
62) 4-Bromofluorobenzene	11.12	95	231498	45.38	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.76%

Target Compounds

					Qvalue
4) Vinyl Chloride	1.39	62	9487	2.509	ug/l 98
21) trans-1,2-Dichloroethene	3.14	96	7815	2.196	ug/l 98
27) cis-1,2-Dichloroethene	4.56	96	345146	82.891	ug/l 89
29) Tetrahydrofuran	5.10	42	1529	0.939	ug/l 93
40) Benzene	6.11	78	7502	0.500	ug/l 99
65) Chlorobenzene	10.12	112	5301	0.535	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019801.D
Acq On : 10 Dec 2020 01:29
Operator : JC/MD
Sample : L4993-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TU032-32-6132-120720

Quant Time: Dec 10 07:03:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

