

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019806.D
 Acq On : 10 Dec 2020 03:25
 Operator : JC/MD
 Sample : L4993-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TU032-32-5777-120820

Quant Time: Dec 10 10:40:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	311935	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	521290	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	473822	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	205386	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	197988	48.29	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.58%
35) Dibromofluoromethane	5.46	113	158199	46.96	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.92%
50) Toluene-d8	8.70	98	628656	48.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.76%
62) 4-Bromofluorobenzene	11.12	95	216396	44.20	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.40%

Target Compounds

					Qvalue
4) Vinyl Chloride	1.39	62	12101	3.417	ug/l 99
21) trans-1,2-Dichloroethene	3.14	96	3487	1.047	ug/l 91
27) cis-1,2-Dichloroethene	4.56	96	10881	2.790	ug/l 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019806.D
Acq On : 10 Dec 2020 03:25
Operator : JC/MD
Sample : L4993-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TU032-32-5777-120820

Quant Time: Dec 10 10:40:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

