

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120920\
 Data File : VX019782.D
 Acq On : 09 Dec 2020 16:34
 Operator : JC/MD
 Sample : L4971-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5S1-B-BOTTOM

Quant Time: Dec 10 06:12:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	334460	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	549316	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	507785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	225684	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	207487	47.20	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.40%		
35) Dibromofluoromethane	5.46	113	169888	47.85	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.70%		
50) Toluene-d8	8.70	98	670600	49.48	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.96%		
62) 4-Bromofluorobenzene	11.12	95	240164	46.56	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.12%		


Target Compounds					Qvalue
16) Acetone	2.42	43	13547	8.199 ug/l	100
20) Methylene Chloride	2.83	84	10944	1.522 ug/l	95
43) Isopropyl Acetate	6.41	43	27743	3.405 ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120920\
Data File : VX019782.D
Acq On : 09 Dec 2020 16:34
Operator : JC/MD
Sample : L4971-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5S1-B-BOTTOM

Quant Time: Dec 10 06:12:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. 0.01 min

Lab File: VX019782.D

Acq: 09 Dec 2020 16:34

Instrument :

MSVOA_X

ClientSampleId :


5S1-B-BOTTOM

Tot Ion:168 Resp: 334460

Ion Ratio Lower Upper

168 100

99 56.4 45.8 68.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63



100000

50000

0

Time-->

5.50 5.60 5.70



#16

Acetone

Concen: 8.199 ug/l

RT: 2.42 min Scan# 218

Delta R.T. 0.00 min

Lab File: VX019782.D


Acq: 09 Dec 2020 16:34

Tgt Ion: 43 Resp: 13547

Ion Ratio Lower Upper

43 100

58 34.5 27.8 41.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.42

8000

6000


4000


2000

0

Time-->

2.35 2.40 2.45 2.50









#20
Methylene Chloride
Concen: 1.522 ug/l
RT: 2.83 min Scan# 286
Delta R.T. 0.00 min
Lab File: VX019782.D
Acq: 09 Dec 2020 16:34




Instrument :
MSVOA_X
ClientSampled :
5S1-B-BOTTOM


Tot Ion:	84	Resp:	10944
Ion	Ratio	Lower	Upper
84	100		
49	118.0	88.6	132.8
51	36.8	27.0	40.6
86	65.2	50.3	75.5



#33
1,2-Dichloroethane-d4
Concen: 47.201 ug/l
RT: 6.03 min Scan# 811
Delta R.T. 0.01 min
Lab File: VX019782.D
Acq: 09 Dec 2020 16:34

Tgt Ion:	65	Resp:	207487
Ion	Ratio	Lower	Upper
65	100		
67	52.8	0.0	105.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. 0.01 min

Lab File: VX019782.D

Acq: 09 Dec 2020 16:34

Instrument :

MSVOA_X

ClientSampleId :

5S1-B-BOTTOM


Tot Ion:114 Resp: 549316

Ion Ratio Lower Upper

114 100

63 20.5 0.0 40.8




88 16.3 0.0 32.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.854 ug/l

RT: 5.46 min Scan# 718

Delta R.T. 0.01 min

Lab File: VX019782.D

Acq: 09 Dec 2020 16:34


Tgt Ion:113 Resp: 169888

Ion Ratio Lower Upper

113 100

111 102.0 81.9 122.9



192 20.7 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.405 ug/l

RT: 6.41 min Scan# 873

Delta R.T. 0.01 min

Lab File: VX019782.D

Acq: 09 Dec 2020 16:34

Instrument :

MSVOA_X

ClientSampled :

5S1-B-BOTTOM





Tot Ion: 43 Resp: 27743

Ion Ratio Lower Upper

43 100

61 21.5 18.2 27.4

87 15.3 12.5 18.7



#50

Toluene-d8

Concen: 49.478 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX019782.D




Acq: 09 Dec 2020 16:34


Tgt Ion: 98 Resp: 670600

Ion Ratio Lower Upper

98 100

100 66.0 51.9 77.9





#62

4-Bromofluorobenzene

Concen: 46.556 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.00 min

Lab File: VX019782.D

Acq: 09 Dec 2020 16:34

Instrument :

MSVOA_X

ClientSampleId :

5S1-B-BOTTOM


Tot Ion: 95 Resp: 240164

Ion Ratio Lower Upper

95 100

174 79.6 0.0 154.6

176 77.4 0.0 155.8






Abundance Ion 94.90 (94.60 to 95.60): VX019713.D (-1638) (-)

Ion 173.80 (173.50 to 174.50): VX019713.D (-1638) (-)

Ion 175.80 (175.50 to 176.50): VX019713.D (-1638) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. 0.01 min

Lab File: VX019782.D

Acq: 09 Dec 2020 16:34


Tgt Ion: 117 Resp: 507785

Ion Ratio Lower Upper

117 100

82 53.8 47.4 71.2

119 32.8 26.6 39.8






Abundance Ion 116.90 (116.60 to 117.60): VX019713.D (-1470) (-)

Ion 82.00 (81.70 to 82.70): VX019713.D (-1470) (-)

Ion 118.90 (118.60 to 119.60): VX019713.D (-1470) (-)

Time-->





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX019782.D

Acq: 09 Dec 2020 16:34

Instrument :

MSVOA_X

ClientSampleId :

5S1-B-BOTTOM


Tot Ion:152 Resp: 225684

Ion Ratio Lower Upper

152 100

115 59.5 41.1 123.3

150 157.7 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

