

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
 Data File : VX025681.D
 Acq On : 09 Dec 2021 23:46
 Operator : JC/MD
 Sample : M5004-08 20X
 Misc : 5.35g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BGKW7

Quant Time: Dec 10 04:31:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 10 04:16:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	116522	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	108185	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	48718	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	36417	45.795	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	91.600%		
7) Chloroethane-d5	1.660	69	29705	45.260	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	90.520%		
11) 1,1-Dichloroethene-d2	2.306	63	50477	39.811	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	79.620%		
21) 2-Butanone-d5	4.458	46	48619	82.133	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	82.130%		
24) Chloroform-d	5.062	84	62400	44.899	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.800%		
26) 1,2-Dichloroethane-d4	5.958	65	41677	48.110	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.220%		
32) Benzene-d6	5.976	84	132090	46.069	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.140%		
36) 1,2-Dichloropropane-d6	7.311	67	38864	43.779	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	87.560%		
41) Toluene-d8	8.653	98	120813	44.641	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	89.280%		
43) trans-1,3-Dichloroprop...	8.951	79	18091	40.534	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	81.060%		
47) 2-Hexanone-d5	9.384	63	34341	74.315	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	74.320%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	52939	42.195	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	84.400%		
66) 1,2-Dichlorobenzene-d4	12.323	152	43483	45.507	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.020%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
Data File : VX025681.D
Acq On : 09 Dec 2021 23:46
Operator : JC/MD
Sample : M5004-08 20X
Misc : 5.35g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BGKW7

Quant Time: Dec 10 04:31:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 10 04:16:38 2021
Response via : Initial Calibration

