

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
 Data File : VX025674.D
 Acq On : 09 Dec 2021 21:03
 Operator : JC/MD
 Sample : M5004-01 20X
 Misc : 5.27g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BGKW0

Quant Time: Dec 10 04:30:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 10 04:16:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	121028	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	110455	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	51364	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	38965	47.175	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.360%
7) Chloroethane-d5	1.660	69	31470	46.164	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.320%
11) 1,1-Dichloroethene-d2	2.306	63	52568	39.916	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.840%
21) 2-Butanone-d5	4.458	46	51427	83.642	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.640%
24) Chloroform-d	5.062	84	66211	45.867	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.740%
26) 1,2-Dichloroethane-d4	5.958	65	43472	48.314	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.620%
32) Benzene-d6	5.976	84	140959	48.152	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.300%
36) 1,2-Dichloropropane-d6	7.311	67	41409	45.687	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.380%
41) Toluene-d8	8.653	98	130203	47.121	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.240%
43) trans-1,3-Dichloroprop...	8.951	79	18707	41.053	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.100%
47) 2-Hexanone-d5	9.384	63	36780	77.957	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.960%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	56092	43.790	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.580%
66) 1,2-Dichlorobenzene-d4	12.323	152	46141	45.801	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
Data File : VX025674.D
Acq On : 09 Dec 2021 21:03
Operator : JC/MD
Sample : M5004-01 20X
Misc : 5.27g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BGKW0

Quant Time: Dec 10 04:30:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 10 04:16:38 2021
Response via : Initial Calibration

