

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
 Data File : VX025680.D
 Acq On : 09 Dec 2021 23:23
 Operator : JC/MD
 Sample : M5004-07ME
 Misc : 6.61g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BGKW6ME

Quant Time: Dec 10 04:31:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 10 04:16:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	114722	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	104803	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	47825	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	36773	46.969	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.940%
7) Chloroethane-d5	1.660	69	29997	46.422	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.840%
11) 1,1-Dichloroethene-d2	2.307	63	50817	40.708	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.420%
21) 2-Butanone-d5	4.459	46	47662	81.780	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.780%
24) Chloroform-d	5.062	84	62817	45.908	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.820%
26) 1,2-Dichloroethane-d4	5.958	65	41747	48.947	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.900%
32) Benzene-d6	5.977	84	132415	47.672	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.340%
36) 1,2-Dichloropropane-d6	7.312	67	39115	45.484	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.960%
41) Toluene-d8	8.653	98	122669	46.789	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.580%
43) trans-1,3-Dichloroprop...	8.952	79	18062	41.775	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.540%
47) 2-Hexanone-d5	9.384	63	33880	75.683	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.680%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	52304	43.035	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.060%
66) 1,2-Dichlorobenzene-d4	12.323	152	43316	46.178	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
Data File : VX025680.D
Acq On : 09 Dec 2021 23:23
Operator : JC/MD
Sample : M5004-07ME
Misc : 6.61g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BGKW6ME

Quant Time: Dec 10 04:31:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 10 04:16:38 2021
Response via : Initial Calibration

