

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
 Data File : VX025686.D
 Acq On : 10 Dec 2021 01:42
 Operator : JC/MD
 Sample : M5004-14 20X
 Misc : 6.88g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BGKX3

Quant Time: Dec 10 04:32:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 10 04:16:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	106168	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	97992	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	43510	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	33532	46.280	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	92.560%		
7) Chloroethane-d5	1.660	69	27757	46.417	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	92.840%		
11) 1,1-Dichloroethene-d2	2.307	63	47136	40.801	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	81.600%		
21) 2-Butanone-d5	4.459	46	45245	83.888	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	83.890%		
24) Chloroform-d	5.062	84	58805	46.439	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.880%		
26) 1,2-Dichloroethane-d4	5.958	65	38694	49.023	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.040%		
32) Benzene-d6	5.977	84	124631	47.989	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.980%		
36) 1,2-Dichloropropane-d6	7.312	67	36167	44.979	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.960%		
41) Toluene-d8	8.653	98	113896	46.462	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	92.920%		
43) trans-1,3-Dichloroprop...	8.952	79	16757	41.450	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	82.900%		
47) 2-Hexanone-d5	9.385	63	32218	76.973	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	76.970%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	49039	43.153	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	86.300%		
66) 1,2-Dichlorobenzene-d4	12.323	152	39859	46.707	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.420%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120921\
Data File : VX025686.D
Acq On : 10 Dec 2021 01:42
Operator : JC/MD
Sample : M5004-14 20X
Misc : 6.88g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BGKX3

Quant Time: Dec 10 04:32:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 10 04:16:38 2021
Response via : Initial Calibration

