

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033340.D
 Acq On : 09 Dec 2022 12:19
 Operator : JC/MD
 Sample : VX1209WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1209WBL01

Quant Time: Dec 09 14:41:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	162007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	255182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48433	46.600	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.200%
35) Dibromofluoromethane	5.391	113	54178	47.203	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.400%
50) Toluene-d8	8.647	98	112052	46.668	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.340%
62) 4-Bromofluorobenzene	11.079	95	86788	47.273	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033340.D
Acq On : 09 Dec 2022 12:19
Operator : JC/MD
Sample : VX1209WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1209WBL01

Quant Time: Dec 09 14:41:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

