

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033343.D
 Acq On : 09 Dec 2022 13:31
 Operator : JC/MD
 Sample : N5944-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR-04-95-105-120722

Quant Time: Dec 09 14:42:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	231068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	42492	45.009	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.020%
35) Dibromofluoromethane	5.391	113	48350	46.522	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.040%
50) Toluene-d8	8.647	98	100188	46.081	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.160%
62) 4-Bromofluorobenzene	11.079	95	78012	46.928	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.860%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.325	96	2606	1.953	ug/l	90
16) Acetone	2.386	43	3099	4.436	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	15862	9.094	ug/l	92
44) Trichloroethene	7.129	130	41667	25.580	ug/l	97
52) Toluene	8.720	92	59573	15.596	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033343.D
Acq On : 09 Dec 2022 13:31
Operator : JC/MD
Sample : N5944-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-95-105-120722

Quant Time: Dec 09 14:42:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

