

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033345.D
 Acq On : 09 Dec 2022 14:17
 Operator : JC/MD
 Sample : N5890-06DL 2X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1827DL

Quant Time: Dec 09 14:42:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243898	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46254	47.311	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.620%
35) Dibromofluoromethane	5.391	113	51509	46.954	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.900%
50) Toluene-d8	8.647	98	110049	47.954	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.900%
62) 4-Bromofluorobenzene	11.079	95	83659	47.677	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.360%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5562	7.688	ug/l	90
52) Toluene	8.720	92	315126	78.158	ug/l	100
64) Tetrachloroethene	9.275	164	2902	1.992	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033345.D
Acq On : 09 Dec 2022 14:17
Operator : JC/MD
Sample : N5890-06DL 2X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1827DL

Quant Time: Dec 09 14:42:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

