

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033351.D
 Acq On : 09 Dec 2022 16:36
 Operator : JC/MD
 Sample : N5966-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-120822

Quant Time: Dec 10 05:57:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	152443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243586	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46991	48.049	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.100%
35) Dibromofluoromethane	5.391	113	51773	47.255	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.520%
50) Toluene-d8	8.647	98	108951	47.537	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.080%
62) 4-Bromofluorobenzene	11.079	95	85092	48.556	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.120%
Target Compounds						
20) Methylene Chloride	2.794	84	530	0.328	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033351.D
Acq On : 09 Dec 2022 16:36
Operator : JC/MD
Sample : N5966-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-120822

Quant Time: Dec 10 05:57:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

