

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033352.D
 Acq On : 09 Dec 2022 16:59
 Operator : JC/MD
 Sample : N5966-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-120822

Quant Time: Dec 10 05:57:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45926	47.913	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.820%
35) Dibromofluoromethane	5.391	113	51374	47.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.900%
50) Toluene-d8	8.646	98	109060	48.152	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.300%
62) 4-Bromofluorobenzene	11.079	95	81826	47.250	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.500%
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	4715	6.648	ug/l	89
20) Methylene Chloride	2.794	84	3281	2.074	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033352.D
Acq On : 09 Dec 2022 16:59
Operator : JC/MD
Sample : N5966-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-120822

Quant Time: Dec 10 05:57:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

