

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033353.D
 Acq On : 09 Dec 2022 17:22
 Operator : JC/MD
 Sample : N5966-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR-04-113-127-120822

Quant Time: Dec 10 05:57:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	155350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46798	46.956	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.920%
35) Dibromofluoromethane	5.385	113	51890	46.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.647	98	110357	47.674	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.340%
62) 4-Bromofluorobenzene	11.079	95	82981	46.883	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.325	96	4453	3.162	ug/l	91
21) trans-1,2-Dichloroethene	3.099	96	671	0.436	ug/l	81
27) cis-1,2-Dichloroethene	4.495	96	30066	16.329	ug/l	88
44) Trichloroethene	7.129	130	19234	11.090	ug/l	98
52) Toluene	8.720	92	85580	21.043	ug/l	100
65) Chlorobenzene	10.079	112	1070	0.246	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	969	0.303	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033353.D
Acq On : 09 Dec 2022 17:22
Operator : JC/MD
Sample : N5966-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-113-127-120822

Quant Time: Dec 10 05:57:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

