

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033356.D
 Acq On : 09 Dec 2022 18:31
 Operator : JC/MD
 Sample : N5976-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-SAMPLE-TANK-1-2-3-4-5

Quant Time: Dec 10 05:57:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144811	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232334	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92350	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47060	50.656	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.320%
35) Dibromofluoromethane	5.385	113	51928	49.692	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.380%
50) Toluene-d8	8.647	98	108697	49.723	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.079	95	84076	50.300	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.600%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033356.D
Acq On : 09 Dec 2022 18:31
Operator : JC/MD
Sample : N5976-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-SAMPLE-TANK-1-2-3-4-5

Quant Time: Dec 10 05:57:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

