

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX120922\  
 Data File : VX033360.D  
 Acq On : 09 Dec 2022 20:03  
 Operator : JC/MD  
 Sample : N5972-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RW9-PT-OUT-02-12082022

Quant Time: Dec 10 05:58:21 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X120722W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 07 12:43:54 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 150298   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 240193   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 209534   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 93353    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 45737    | 47.434   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 94.860%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 50425    | 46.675   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 93.360%  |
| 50) Toluene-d8              | 8.647  | 98    | 107081   | 47.381   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 94.760%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 79825    | 46.194   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 92.380%  |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX120922\  
Data File : VX033360.D  
Acq On : 09 Dec 2022 20:03  
Operator : JC/MD  
Sample : N5972-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RW9-PT-OUT-02-12082022

Quant Time: Dec 10 05:58:21 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X120722W.M  
Quant Title : SW846 8260  
QLast Update : Wed Dec 07 12:43:54 2022  
Response via : Initial Calibration

