

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
 Data File : VX033355.D
 Acq On : 09 Dec 2022 18:08
 Operator : JC/MD
 Sample : N5966-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR-04-128-142-120822

Quant Time: Dec 10 05:57:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	145511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232911	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	44459	47.626	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.260%
35) Dibromofluoromethane	5.385	113	49762	47.501	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.000%
50) Toluene-d8	8.647	98	105268	48.035	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.060%
62) 4-Bromofluorobenzene	11.079	95	79832	47.642	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.280%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.325	96	6202	4.701	ug/l	92
21) trans-1,2-Dichloroethene	3.099	96	431	0.299	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	22610	13.110	ug/l	90
44) Trichloroethene	7.123	130	13617	8.294	ug/l	99
52) Toluene	8.720	92	132748	34.477	ug/l	100
65) Chlorobenzene	10.079	112	977	0.238	ug/l #	76
91) 1,2-Dichlorobenzene	12.335	146	981	0.325	ug/l	84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120922\
Data File : VX033355.D
Acq On : 09 Dec 2022 18:08
Operator : JC/MD
Sample : N5966-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-128-142-120822

Quant Time: Dec 10 05:57:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08




Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-128-142-120822


Tgt Ion:168 Resp: 145511
Ion Ratio Lower Upper
168 100
99 59.3 46.1 69.1



#12
1,1-Dichloroethene
Concen: 4.701 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08

Tgt Ion: 96 Resp: 6202
Ion Ratio Lower Upper
96 100
61 143.2 125.1 187.7
98 66.6 50.6 76.0









#21
trans-1,2-Dichloroethene
Concen: 0.299 ug/l
RT: 3.099 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08




Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-128-142-120822


Tgt Ion: 96 Resp: 431
Ion Ratio Lower Upper
96 100
61 141.0 106.6 160.0
98 59.3 51.0 76.6



#27
cis-1,2-Dichloroethene
Concen: 13.110 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.000 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08

Tgt Ion: 96 Resp: 22610
Ion Ratio Lower Upper
96 100
61 126.1 0.0 286.0
98 67.3 0.0 129.6





#33

1,2-Dichloroethane-d4

Concen: 47.626 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX033355.D

Acq: 09 Dec 2022 18:08

Instrument :

MSVOA_X

ClientSampleId :





GW-BR-04-128-142-120822

Tgt Ion: 65 Resp: 44459

Ion Ratio Lower Upper

65 100

67 50.8 0.0 101.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX033355.D

Acq: 09 Dec 2022 18:08




Tgt Ion: 114 Resp: 232911


Ion Ratio Lower Upper

114 100

63 19.7 0.0 41.6

88 16.0 0.0 33.4





#35

Dibromofluoromethane

Concen: 47.501 ug/l

RT: 5.385 min Scan# 706

Delta R.T. -0.006 min

Lab File: VX033355.D


Acq: 09 Dec 2022 18:08

Instrument :

MSVOA_X

ClientSampleId :

GW-BR-04-128-142-120822






Tgt Ion:113 Resp: 49762

Ion Ratio Lower Upper

113 100

111 103.6 82.4 123.6

192 18.9 15.4 23.2



#44

Trichloroethene

Concen: 8.294 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.006 min

Lab File: VX033355.D




Acq: 09 Dec 2022 18:08


Tgt Ion:130 Resp: 13617

Ion Ratio Lower Upper

130 100

95 98.5 0.0 195.2





#50

Toluene-d8

Concen: 48.035 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX033355.D

Acq: 09 Dec 2022 18:08

Instrument :

MSVOA_X

ClientSampleId :





GW-BR-04-128-142-120822

Tgt Ion: 98 Resp: 105268

Ion Ratio Lower Upper

98 100

100 63.5 50.9 76.3



#52

Toluene

Concen: 34.477 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. 0.000 min

Lab File: VX033355.D




Acq: 09 Dec 2022 18:08


Tgt Ion: 92 Resp: 132748

Ion Ratio Lower Upper

92 100

91 173.2 138.3 207.5









#62
4-Bromofluorobenzene
Concen: 47.642 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08




Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-128-142-120822


Tgt Ion: 95 Resp: 79832
Ion Ratio Lower Upper
95 100
174 79.8 0.0 164.0
176 77.8 0.0 154.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08

Tgt Ion: 117 Resp: 202910
Ion Ratio Lower Upper
117 100
82 53.3 44.3 66.5
119 31.5 25.7 38.5









#65
Chlorobenzene
Concen: 0.238 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08




Instrument : MSVOA_X
ClientSampleId : GW-BR-04-128-142-120822


Tgt Ion:112 Resp: 977
Ion Ratio Lower Upper
112 100
114 18.2 25.4 38.2#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX033355.D
Acq: 09 Dec 2022 18:08

Tgt Ion:152 Resp: 90999
Ion Ratio Lower Upper
152 100
115 60.2 44.0 132.0
150 158.3 0.0 351.8





#91
 1,2-Dichlorobenzene
 Concen: 0.325 ug/l
 RT: 12.335 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX033355.D
 Acq: 09 Dec 2022 18:08

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR-04-128-142-120822

Tgt Ion:	146	Resp:	981
Ion	Ratio	Lower	Upper
146	100		
111	36.0	21.5	64.5
148	48.4	32.0	96.2

