

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
 Data File : VX044176.D
 Acq On : 09 Dec 2024 10:35
 Operator : JC/MD
 Sample : VX1209WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1209WBL01

Quant Time: Dec 10 01:36:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	127252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	245969	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	212379	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	114252	52.347	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.700%		
35) Dibromofluoromethane	5.379	113	81564	46.407	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.820%		
50) Toluene-d8	8.647	98	303349	50.070	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.140%		
62) 4-Bromofluorobenzene	11.079	95	101473	49.214	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 98.420%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
Data File : VX044176.D
Acq On : 09 Dec 2024 10:35
Operator : JC/MD
Sample : VX1209WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1209WBL01

Quant Time: Dec 10 01:36:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

