

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
 Data File : VX044179.D
 Acq On : 09 Dec 2024 11:47
 Operator : JC/MD
 Sample : P5098-01
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-04-12042024

Quant Time: Dec 09 12:13:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	107295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	205040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77672	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94127	51.148	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.300%
35) Dibromofluoromethane	5.379	113	63517	43.353	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.700%
50) Toluene-d8	8.647	98	251605	49.819	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	85714	49.869	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
Data File : VX044179.D
Acq On : 09 Dec 2024 11:47
Operator : JC/MD
Sample : P5098-01
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-04-12042024

Quant Time: Dec 09 12:13:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

