

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
 Data File : VX044185.D
 Acq On : 09 Dec 2024 14:07
 Operator : JC/MD
 Sample : P5164-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20241206

Quant Time: Dec 10 01:42:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	113714	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	212837	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99361	50.944	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.880%
35) Dibromofluoromethane	5.385	113	70753	46.523	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.040%
50) Toluene-d8	8.647	98	259184	49.440	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	88889	49.822	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.640%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.318	101	530	0.385	ug/l	92
16) Acetone	2.386	43	629	0.702	ug/l #	87
44) Trichloroethene	7.122	130	54650	32.703	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
Data File : VX044185.D
Acq On : 09 Dec 2024 14:07
Operator : JC/MD
Sample : P5164-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D2-20241206

Quant Time: Dec 10 01:42:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

