

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
 Data File : VX044186.D
 Acq On : 09 Dec 2024 14:30
 Operator : JC/MD
 Sample : P5164-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-20241206

Quant Time: Dec 10 01:42:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105637	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	204810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96390	53.200	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.400%
35) Dibromofluoromethane	5.385	113	65368	44.666	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.340%
50) Toluene-d8	8.647	98	251001	49.755	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.520%
62) 4-Bromofluorobenzene	11.079	95	85098	49.566	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.140%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2612	3.139	ug/l #	89
20) Methylene Chloride	2.776	84	630	0.446	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
Data File : VX044186.D
Acq On : 09 Dec 2024 14:30
Operator : JC/MD
Sample : P5164-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-20241206

Quant Time: Dec 10 01:42:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

