

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
 Data File : VX044188.D
 Acq On : 09 Dec 2024 15:17
 Operator : JC/MD
 Sample : P5164-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20241206

Quant Time: Dec 10 01:43:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	120459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	226191	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104646	50.650	ug/l	0.01
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.300%			
35) Dibromofluoromethane	5.385	113	75400	46.651	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.300%			
50) Toluene-d8	8.646	98	275938	49.528	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.060%			
62) 4-Bromofluorobenzene	11.079	95	88621	46.739	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.480%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.312	101	478	0.328	ug/l #	76
16) Acetone	2.385	43	1505	1.586	ug/l	99
44) Trichloroethene	7.122	130	43452	24.467	ug/l	99
64) Tetrachloroethene	9.274	164	2879	2.264	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
Data File : VX044188.D
Acq On : 09 Dec 2024 15:17
Operator : JC/MD
Sample : P5164-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20241206

Quant Time: Dec 10 01:43:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

