

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
 Data File : VX044193.D
 Acq On : 09 Dec 2024 17:14
 Operator : JC/MD
 Sample : P5164-04 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20241206

Quant Time: Dec 10 01:46:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109323	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	207584	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182923	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82399	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94710	50.510	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.020%
35) Dibromofluoromethane	5.379	113	68391	46.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.220%
50) Toluene-d8	8.647	98	252866	49.455	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.079	95	89579	51.479	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.960%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.313	101	1244	0.941	ug/l	96
44) Trichloroethene	7.123	130	129632	79.535	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120924\
Data File : VX044193.D
Acq On : 09 Dec 2024 17:14
Operator : JC/MD
Sample : P5164-04 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D2-20241206

Quant Time: Dec 10 01:46:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

