

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120925\
 Data File : VX048786.D
 Acq On : 09 Dec 2025 15:49
 Operator : JC/MD
 Sample : Q3803-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J

Quant Time: Dec 10 00:26:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	207524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	423652	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	448489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	242407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	148729	53.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.800%
35) Dibromofluoromethane	5.373	113	135240	46.363	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.720%
50) Toluene-d8	8.635	98	484910	47.424	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.840%
62) 4-Bromofluorobenzene	11.061	95	216901	61.516	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	123.040%#
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.941	59	5315	11.819	ug/l	99
13) Acrolein	2.240	56	722	41.573	ug/l #	1
15) Acrylonitrile	3.075	53	606	0.432	ug/l #	77
16) Acetone	2.374	43	98769	92.629	ug/l	100
18) Methyl Acetate	2.703	43	5534	1.845	ug/l	95
20) Methylene Chloride	2.794	84	124447	46.693	ug/l	96
25) 2-Butanone	4.544	43	18421	11.705	ug/l	99
29) Tetrahydrofuran	5.026	42	514	0.428	ug/l #	40
31) Cyclohexane	5.471	56	1558	0.414	ug/l #	78
36) 1,1-Dichloropropene	5.538	75	15774	3.948	ug/l #	49
37) Ethyl Acetate	4.544	43	18421	3.859	ug/l #	71
38) Carbon Tetrachloride	5.538	117	21053	4.652	ug/l #	13
39) Methylcyclohexane	7.361	83	1365	0.289	ug/l #	88
40) Benzene	6.025	78	1396428	117.904	ug/l	99
42) 1,2-Dichloroethane	6.025	62	11427	2.726	ug/l #	73
43) Isopropyl Acetate	6.324	43	14503	2.006	ug/l	96
48) Methyl methacrylate	7.781	41	54326	14.940	ug/l #	41
51) 4-Methyl-2-Pentanone	8.555	43	341291	76.064	ug/l	91
52) Toluene	8.702	92	536815	75.164	ug/l	100
54) cis-1,3-Dichloropropene	8.531	75	21626	4.519	ug/l #	64
55) 1,1,2-Trichloroethane	9.000	97	3145	1.094	ug/l #	19
57) 1,3-Dichloropropane	9.464	76	1820	0.381	ug/l #	42
59) 2-Hexanone	9.421	43	2397	0.761	ug/l	95
67) Ethyl Benzene	10.177	91	55086	3.439	ug/l	98
68) m/p-Xylenes	10.281	106	106549	17.075	ug/l	98
69) o-Xylene	10.622	106	73291	12.534	ug/l	96
70) Styrene	10.640	104	24268	2.418	ug/l #	63
74) N-amyl acetate	10.842	43	1605	0.226	ug/l #	26
76) 1,2,3-Trichloropropane	11.061	75	107880	23.894	ug/l #	30
78) n-propylbenzene	11.287	91	4326	0.229	ug/l	90
79) 2-Chlorotoluene	11.360	91	2832	0.245	ug/l #	50
80) 1,3,5-Trimethylbenzene	11.433	105	11899	0.886	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.061	75	107880	56.758	ug/l #	3
84) 1,2,4-Trimethylbenzene	11.732	105	63103	4.659	ug/l	70
85) sec-Butylbenzene	11.732	105	63103	3.732	ug/l #	57
95) Naphthalene	13.756	128	361957	26.695	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120925\
Data File : VX048786.D
Acq On : 09 Dec 2025 15:49
Operator : JC/MD
Sample : Q3803-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHASE-J

Quant Time: Dec 10 00:26:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120925\
 Data File : VX048786.D
 Acq On : 09 Dec 2025 15:49
 Operator : JC/MD
 Sample : Q3803-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J

Quant Time: Dec 10 00:26:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration









#1
 Pentafluorobenzene
 Concen: 50.000 ug/l
 RT: 5.544 min Scan# 71
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J


Tgt Ion:168 Resp: 207524
 Ion Ratio Lower Upper
 168 100
 99 55.0 44.2 66.4



#11
 Tert butyl alcohol
 Concen: 11.819 ug/l
 RT: 2.941 min Scan# 304
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 59 Resp: 5315
 Ion Ratio Lower Upper
 59 100
 57 9.7 8.1 12.1









#13
 Acrolein
 Concen: 41.573 ug/l
 RT: 2.240 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J


Tgt Ion: 56 Resp: 722
 Ion Ratio Lower Upper
 56 100
 55 155.3 55.8 83.6#



#15
 Acrylonitrile
 Concen: 0.432 ug/l
 RT: 3.075 min Scan# 326
 Delta R.T. 0.018 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 53 Resp: 606
 Ion Ratio Lower Upper
 53 100
 52 76.9 66.3 99.5
 51 0.0 28.7 43.1#









#16
 Acetone
 Concen: 92.629 ug/l
 RT: 2.374 min Scan# 211
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument : MSVOA_X
 Client SampleID : CHASE-J


Tgt Ion: 43 Resp: 98769
 Ion Ratio Lower Upper
 43 100
 58 31.1 25.0 37.4



#18
 Methyl Acetate
 Concen: 1.845 ug/l
 RT: 2.703 min Scan# 265
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49


Tgt Ion: 43 Resp: 5534
 Ion Ratio Lower Upper
 43 100
 74 26.7 19.2 28.8








#20
 Methylene Chloride
 Concen: 46.693 ug/l
 RT: 2.794 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J




Tgt Ion: 84 Resp: 124447

Ion	Ratio	Lower	Upper
84	100		
49	116.7	98.7	148.1
51	36.1	29.4	44.2
86	64.2	52.5	78.7






#25
 2-Butanone
 Concen: 11.705 ug/l
 RT: 4.544 min Scan# 567
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49



Tgt Ion: 43 Resp: 18421


Ion	Ratio	Lower	Upper
43	100		
72	28.9	22.6	34.0








#29
 Tetrahydrofuran
 Concen: 0.428 ug/l
 RT: 5.026 min Scan# 646
 Delta R.T. 0.043 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J



Tgt Ion: 42 Resp: 514




Ion	Ratio	Lower	Upper
42	100		
72	10.1	39.4	59.2#
71	3.9	35.8	53.6#




#31
 Cyclohexane
 Concen: 0.414 ug/l
 RT: 5.471 min Scan# 719
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 56 Resp: 1558


Ion	Ratio	Lower	Upper
56	100		
69	38.8	24.2	36.2#
84	67.6	72.1	108.1#








#33
 1,2-Dichloroethane-d4
 Concen: 53.397 ug/l
 RT: 5.940 min Scan# 796
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49


Instrument : MSVOA_X
 ClientSampleId : CHASE-J





Tgt Ion: 65 Resp: 148729
 Ion Ratio Lower Upper
 65 100
 67 52.9 0.0 107.4




#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.745 min Scan# 928
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Tgt Ion: 114 Resp: 423652
 Ion Ratio Lower Upper
 114 100
 63 20.8 0.0 39.0
 88 16.1 0.0 31.8








#35
 Dibromofluoromethane
 Concen: 46.363 ug/l
 RT: 5.373 min Scan# 703
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49


Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J





Tgt Ion:113 Resp: 135240
 Ion Ratio Lower Upper
 113 100
 111 101.6 79.5 119.3
 192 19.7 16.1 24.1




#36
 1,1-Dichloropropene
 Concen: 3.948 ug/l
 RT: 5.538 min Scan# 730
 Delta R.T. -0.146 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49



Tgt Ion: 75 Resp: 15774
 Ion Ratio Lower Upper
 75 100
 110 9.3 18.6 55.8#
 77 0.0 25.3 37.9#









#37
 Ethyl Acetate
 Concen: 3.859 ug/l
 RT: 4.544 min Scan# 509
 Delta R.T. -0.152 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Instrument :
 MSVOA_X
 Client SampleId :
 CHASE-J

Tgt Ion: 43 Resp: 18421




Ion	Ratio	Lower	Upper
43	100		
61	0.0	9.7	14.5#
70	0.0	8.2	12.4#




#38
 Carbon Tetrachloride
 Concen: 4.652 ug/l
 RT: 5.538 min Scan# 730
 Delta R.T. -0.128 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 117 Resp: 21053


Ion	Ratio	Lower	Upper
117	100		
119	4.4	80.5	120.7#
121	0.0	25.4	38.0#








#39
 Methylcyclohexane
 Concen: 0.289 ug/l
 RT: 7.361 min Scan# 10
 Delta R.T. -0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49


Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J





Tgt Ion: 83 Resp: 1365
 Ion Ratio Lower Upper
 83 100
 55 92.9 60.6 90.8#
 98 49.4 39.3 58.9




#40
 Benzene
 Concen: 117.904 ug/l
 RT: 6.025 min Scan# 810
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49



Tgt Ion: 78 Resp: 1396428
 Ion Ratio Lower Upper
 78 100
 77 23.9 18.8 28.2









#42
 1,2-Dichloroethane
 Concen: 2.726 ug/l
 RT: 6.025 min Scan# 810
 Delta R.T. -0.049 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument : MSVOA_X
 ClientSampleId : CHASE-J


Tgt Ion: 62 Resp: 11427
 Ion Ratio Lower Upper
 62 100
 98 0.0 0.0 20.2



#43
 Isopropyl Acetate
 Concen: 2.006 ug/l
 RT: 6.324 min Scan# 859
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 43 Resp: 14503
 Ion Ratio Lower Upper
 43 100
 61 19.6 16.8 25.2
 87 12.3 11.8 17.6









#48
 Methyl methacrylate
 Concen: 14.940 ug/l
 RT: 7.781 min Scan# 1098
 Delta R.T. 0.104 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Instrument : MSVOA_X
 ClientSampleId : CHASE-J

Tgt Ion: 41 Resp: 54326




Ion	Ratio	Lower	Upper
41	100		
69	0.0	68.5	102.7#
39	51.1	43.5	65.3




#50
 Toluene-d8
 Concen: 47.424 ug/l
 RT: 8.635 min Scan# 1238
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 98 Resp: 484910

Ion	Ratio	Lower	Upper
98	100		
100	65.5	53.4	80.0









#51
 4-Methyl-2-Pentanone
 Concen: 76.064 ug/l
 RT: 8.555 min Scan# 1225
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument : MSVOA_X
 ClientSampleId : CHASE-J


Tgt Ion: 43 Resp: 341291
 Ion Ratio Lower Upper
 43 100
 58 34.5 32.2 48.4



#52
 Toluene
 Concen: 75.164 ug/l
 RT: 8.702 min Scan# 1249
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49


Tgt Ion: 92 Resp: 536815
 Ion Ratio Lower Upper
 92 100
 91 170.8 136.3 204.5








#54
 cis-1,3-Dichloropropene
 Concen: 4.519 ug/l
 RT: 8.531 min Scan# 1191
 Delta R.T. 0.183 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Instrument : MSVOA_X
 ClientSampleId : CHASE-J




Tgt Ion: 75 Resp: 21626

Ion	Ratio	Lower	Upper
75	100		
77	0.8	26.1	39.1#
39	33.3	38.2	57.4#






#55
 1,1,2-Trichloroethane
 Concen: 1.094 ug/l
 RT: 9.000 min Scan# 1298
 Delta R.T. -0.134 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49



Tgt Ion: 97 Resp: 3145

Ion	Ratio	Lower	Upper
97	100		
83	0.0	66.0	99.0#
85	0.0	43.2	64.8#
99	7.1	50.0	75.0#









#57
 1,3-Dichloropropane
 Concen: 0.381 ug/l
 RT: 9.464 min Scan# 1374
 Delta R.T. 0.171 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J


Tgt Ion: 76 Resp: 1820
 Ion Ratio Lower Upper
 76 100
 78 0.0 26.0 39.0#



#59
 2-Hexanone
 Concen: 0.761 ug/l
 RT: 9.421 min Scan# 1367
 Delta R.T. 0.006 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 43 Resp: 2397
 Ion Ratio Lower Upper
 43 100
 58 51.6 27.8 83.4









#62
 4-Bromofluorobenzene
 Concen: 61.516 ug/l
 RT: 11.061 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J

Tgt Ion: 95 Resp: 216901




Ion	Ratio	Lower	Upper
95	100		
174	78.6	0.0	157.8
176	76.9	0.0	154.0




#63
 Chlorobenzene-d5
 Concen: 50.000 ug/l
 RT: 10.037 min Scan# 1468
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 117 Resp: 448489

Ion	Ratio	Lower	Upper
117	100		
82	54.9	44.1	66.1
119	32.3	25.2	37.8









#67
 Ethyl Benzene
 Concen: 3.439 ug/l
 RT: 10.177 min Scan# 1491
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument : MSVOA_X
 ClientSampleId : CHASE-J


Tgt Ion: 91 Resp: 55086
 Ion Ratio Lower Upper
 91 100
 106 30.2 25.3 37.9



#68
 m/p-Xylenes
 Concen: 17.075 ug/l
 RT: 10.281 min Scan# 1508
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 106 Resp: 106549
 Ion Ratio Lower Upper
 106 100
 91 202.4 164.2 246.4









#69
 o-Xylene
 Concen: 12.534 ug/l
 RT: 10.622 min Scan# 1564
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument : MSVOA_X
 ClientSampleId : CHASE-J


Tgt Ion:106 Resp: 73291
 Ion Ratio Lower Upper
 106 100
 91 210.6 108.1 324.3



#70
 Styrene
 Concen: 2.418 ug/l
 RT: 10.640 min Scan# 1567
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49


Tgt Ion:104 Resp: 24268
 Ion Ratio Lower Upper
 104 100
 78 78.9 40.6 60.8#
 103 79.1 43.4 65.2#








#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.006 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49


Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J





Tgt Ion:152 Resp: 242407
 Ion Ratio Lower Upper
 152 100
 115 57.9 42.1 126.4
 150 157.6 0.0 347.8




#74
 N-ethyl acetate
 Concen: 0.226 ug/l
 RT: 10.842 min Scan# 1600
 Delta R.T. 0.018 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49



Tgt Ion: 43 Resp: 1605
 Ion Ratio Lower Upper
 43 100
 70 102.6 33.9 50.9#
 55 54.6 19.1 28.7#
 61 0.0 19.0 28.6#









#76
 1,2,3-Trichloropropane
 Concen: 23.894 ug/l
 RT: 11.061 min Scan# 1636
 Delta R.T. -0.158 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J


Tgt Ion: 75 Resp: 107880
 Ion Ratio Lower Upper
 75 100
 77 1.1 24.3 73.0#



#78
 n-propylbenzene
 Concen: 0.229 ug/l
 RT: 11.287 min Scan# 1673
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 91 Resp: 4326
 Ion Ratio Lower Upper
 91 100
 120 18.0 11.4 34.2









#79
 2-Chlorotoluene
 Concen: 0.245 ug/l
 RT: 11.360 min Scan# 1685
 Delta R.T. 0.012 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J


Tgt Ion: 91 Resp: 2832
 Ion Ratio Lower Upper
 91 100
 126 5.3 16.8 50.3#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.886 ug/l
 RT: 11.433 min Scan# 1697
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 105 Resp: 11899
 Ion Ratio Lower Upper
 105 100
 120 47.3 24.4 73.2









#81
 trans-1,4-Dichloro-2-butene
 Concen: 56.758 ug/l
 RT: 11.061 min Scan# 1
 Delta R.T. 0.061 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49




Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-J


Tgt Ion: 75 Resp: 107880
 Ion Ratio Lower Upper
 75 100
 53 0.0 79.9 119.9#
 89 0.0 67.1 100.7#



#84
 1,2,4-Trimethylbenzene
 Concen: 4.659 ug/l
 RT: 11.732 min Scan# 1746
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion: 105 Resp: 63103
 Ion Ratio Lower Upper
 105 100
 120 25.0 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 3.732 ug/l
 RT: 11.732 min Scan# 11
 Delta R.T. -0.140 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Instrument : MSVOA_X
 ClientSampleId : CHASE-J

Tgt Ion:105 Resp: 63103
 Ion Ratio Lower Upper
 105 100
 134 0.0 9.8 29.5#



#95
 Naphthalene
 Concen: 26.695 ug/l
 RT: 13.756 min Scan# 2078
 Delta R.T. 0.000 min
 Lab File: VX048786.D
 Acq: 09 Dec 2025 15:49

Tgt Ion:128 Resp: 361957
 Ion Ratio Lower Upper
 128 100
 127 12.6 10.2 15.2
 129 10.7 8.6 13.0

